

VERIFICATION OF CONFORMITY

No. : PTC23070307901S-LD01
Applicant : Dongguan Habotest Instrument Technology Co., Ltd.
Address : No.8, 4th Qinghu Xingye Road, Qingxi Town, Dongguan City, Guangdong Province, China
Manufacturer : Dongguan Habotest Instrument Technology Co., Ltd.
Address : No.8, 4th Qinghu Xingye Road, Qingxi Town, Dongguan City, Guangdong Province, China
Factory : Dongguan Habotest Instrument Technology Co., Ltd.
Address : No.8, 4th Qinghu Xingye Road, Qingxi Town, Dongguan City, Guangdong Province, China
Product : DIGITAL INSULATION TESTER
Trade Mark : HABOTEST
Model : HT5203, HT5205
Ratings : Input:DC 1.5Vx8 AA battery; 1000V CAT III; 600V CAT IV; Class II

The submitted sample of the above equipment has been tested and found to comply with the following European Directive:

Low voltage (LVD) Directive - 2014/35/EU

The standard(s) used for showing compliance with the essential requirements:

Applicable Standard(s)	Test Report(s) Number
EN 61010-1:2010/A1:2019/AC:2019-04 EN IEC 61010-2-030:2021 EN IEC 61557-2:2021	PTC23070307901S-LD01

This verification is part of the full test report(s) and should be read in conjunction with it. This verification is based on an evaluation of one sample of above mentioned product. It does not imply assessment of the production of the product. Without the written approval of Precise Testing & Certification (Guangdong) Co., Ltd., this verification is not permitted to be reproduced, except in full. It is not permitted to use the test lab's logo. The CE marking may only be used if all the relevant and effective European Directives are applicable.



Jacky Ou

Jacky Ou
Manager
Date: 2023-07-14

